CERTIFICATE OF CONFORMITY

Order code: **800-951**

Pipe Surface Thermomter

Ø60mm s/s Dial with Ø50mm Dial Face

Range: 0 to 120°C or 32 to 250°F

Resolution: 2°C/4°F

Accuracy: ±2°C

Made in China

Conforms to REACH Legislation

Also conforms to the following regulations; .

DECLARATION OF CONFORMITY

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

Where appropriate and required, the product complies with EEC Directives with regard to European Legislation and CE Marking requirements.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd



Andy Reid

Quality Manager

Valid as of date of purchase.

Easting Close
Worthing
West Sussex
BN14 8HQ
T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com